



Government of West Bengal
Office of the Principal
Calcutta National Medical College
32, Gorachand Road, Kolkata-14

Memo No. CNMC/ 285 /P

Date: 02/02/2021

TENDER NOTICE

Principal, Calcutta National Medical College, Kolkata invites Sealed Tender from bonafied suppliers for the supply of "A Scan Ocular Biometry Machine" at the Department of Ophthalmology under Calcutta National Medical College, Kolkata. Each tender should contain two envelopes marked as "A" & "B" superscripting "Tender for A Scan Ocular Biometry" with Tender Notice No & Date & Name of the Bidder. Envelope "A" should contain the documents as technical bid. The Envelope "B" should contain price bid in duplicate.

The tender paper will have to be submitted on or before 12.02.2021 by 3.00 pm at receiving section. The sealed tender, Envelope "A" will be opened on 12.02.2021 at 3.30 pm at the chamber of the Principal, Calcutta National Medical College in presence of the intending tender bidders. The date of opening the price bid (Envelope B) will be notified later on. The Undersigned reserves the right to reject or accept any offer partly or wholly without assigning any reason thereof. The decision of the Undersigned in all respects is final and binding upon all concern.

General Terms & Conditions

1. Demand Draft in original showing deposit of Rs. 10,000.00/- (Rupees five thousands only) as Earnest Money in favour of the undersigned. Without demand draft, the tender will be declared as invalid.
2. Attested Current Trade License, Income Tax Return submitted in last financial year, PAN Card, Professional Tax Registration certificate & GST Certificate (if applicable) must be attached with the tender, without such documents the tender will be declared as invalid.
3. Copy of evidence that the firm is acquainted/ equipped with the technical know-how of the equipment.
4. In case of failure to meet the technical specification of the equipment in the technical bid, the price bid will not be considered.
5. The bidder should duly sign all the pages of bidding documents with seal while submitting the technical part of the tender, i.e., Envelope A
6. Unsigned offers submitted by the bidder will be considered invalid.
7. Rates should be quoted both in figure & in words clearly (Indian Rupees) including all types of charges & taxes i.e., at the ultimate destination of the equipment. Price should be firm till execution of order.
8. The tenderers must quote basic rate along with 3 years comprehensive onsite warranty of entire system (including all spares & labour) from the date of completion of the satisfactory installation.
9. The lowest bidder must have to be given the physical demonstration as per specification after opening the tender as well as the price bid. Failure to provide demonstration within 3 days, the price bid will be treated cancelled and invalid. Under this circumstances, 2nd lowest bidder will be asked to provide physical demonstration in this regard
10. Approved tenderer is required to submit 10(ten) percent security deposit of the net value of the item at the time of delivery of the equipment.
11. In case of refusal/failure to supply, the EMD of the selected tenderer will be forfeited.
12. Packing, forwarding & installation charges should be borne by the supplier. Rate should be inclusive of all.
13. Time of completion of delivery will be directed in order.
14. Service back up of the quoted equipment/instrument should be available in Kolkata.
15. Earnest Money of the unsuccessful bidders will be refunded in due course.

16. The Suppliers should have knowledge to supply the same kind of equipments to other Medical Colleges within the jurisdiction of Kolkata.
17. All suits arising out of the tender/ enquiry and subsequent purchase order/letter of acceptance (if any) are subjected to the jurisdiction in Kolkata only.

Enclosed:

Technical Specifications of "A Scan Ocular Biometry Machine"

July
02.02.2021

[Signature]
02/02/2021

Principal
Calcutta National Medical College

Date: 02/02/2021

Memo No. CNMC/ 285 /P

Copy forwarded for information & necessary publication to:

1. The DDHS (E & S), Central Medical Stores,
2. The MSVP, CNMC&H
3. The Accounts Officer, CNMC
4. I.T. Cell, Swasthya Bhavan, for publication in the website
5. Notice Board, CNMC, Kolkata-14
6. CNMC, website

July
02.02.2021

[Signature]
02/02/2021

Principal
Calcutta National Medical College

A-SCAN: (A-SCAN WITH/ WITHOUT ULTRASOUND PACHYMETRY)

182614/2018/NHM SEC(H&FW)

1. Probe: 10MHz with fixation red light
2. Resolution: 0.01 mm
3. Biometry accuracy: ± 0.06 mm
4. Gain: ≥ 98 dB with adjustable range of 0-59 dB
5. Measuring range: 15-39 mm
6. Measuring parameters: Anterior chamber depth, Lens Thickness, Vitreous length and Axial Length
7. Measuring modes: Automatic (for Normal, Dense cataract, Aphakia and Pseudophakia); Manual
8. 8 groups of standard deviation
9. IOL Power calculations: SRK/T, SRK-II, BINKHOST-II, HOLLADAY, HOFFER-Q and HAIGIS
10. Save upto 50 groups of IOL calculation results and 4 groups of IOL constants

(Signature)
19/4/16
Director
Regional Institute of Ophthalmology
Medical College, Kolkata-700 072

(Signature)
Biswajit Bhattacharya

(Signature)
2/2/2014

(Signature)
19/4/16