



Government of West Bengal
Office of the Principal
Calcutta National Medical College
32, Gorachand Road, Kolkata-14

Memo No. CNMC/ 286 /P

Date: 02/02/2021

TENDER NOTICE

Principal, Calcutta National Medical College, Kolkata invites Sealed Tender from bonafied suppliers for the supply of "Medical Equipment for the department of Ophthalmology" under Calcutta National Medical College, Kolkata. Each tender should contain two envelopes marked as "A" & "B" superscripting "Tender for Medical Equipment of the department of Ophthalmology" with Tender Notice No & Date & Name of the Bidder. Envelope "A" should contain the documents as technical bid. The Envelope "B" should contain price bid in duplicate. The items are as follows,

Serial No	Name of the items	Quantity
1	Applanation Tonometer	2
2	90 Dioptre Lense	3
3	Two Mirror Gonioscope	2
4	Zeiss Four Mirror Gonioscope	1
5	Reister Schiotz indentation tonometer	3
6	Indirect Ophthalmoscope	2

The tender paper will have to be submitted on or before 15.02.2021 by 3.00 pm at receiving section. The sealed tender, Envelope "A" will be opened on 15.02.2021 at 3.30 pm at the chamber of the Principal, Calcutta National Medical College in presence of the intending tender bidders. The date of opening the price bid (Envelope B) will be notified later on. The Undersigned reserves the right to reject or accept any offer partly or wholly without assigning any reason thereof. The decision of the Undersigned in all respects is final and binding upon all concern.

General Terms & Conditions

1. Demand Draft in original showing deposit of Rs. 10,000.00/- (Rupees five thousands only) as Earnest Money in favour of the undersigned. Without demand draft, the tender will be declared as invalid.
2. Attested Current Trade License, Income Tax Return submitted in last financial year, PAN Card, Professional Tax Registration certificate & GST Certificate (if applicable) must be attached with the tender, without such documents the tender will be declared as invalid.
3. Copy of evidence that the firm is acquainted/ equipped with the technical know-how of the equipment.
4. In case of failure to meet the technical specification of the equipment in the technical bid, the price bid will not be considered.
5. The bidder should duly sign all the pages of bidding documents with seal while submitting the technical part of the tender, i.e., Envelope A
6. Unsigned offers submitted by the bidder will be considered invalid.
7. Rates should be quoted both in figure & in words clearly (Indian Rupees) including all types of charges & taxes i.e., at the ultimate destination of the equipment. Price should be firm till execution of order.
8. The tenderers must quote basic rate along with 3 years comprehensive onsite warranty of entire system (including all spares & labour) from the date of completion of the satisfactory installation.

9. The lowest bidder must have to be given the physical demonstration as per specification after opening the tender as well as the price bid. Failure to provide demonstration within 3 days, the price bid will be treated cancelled and invalid. Under this circumstances, 2nd lowest bidder will be asked to provide physical demonstration in this regard
10. Approved tenderer is required to submit 10(ten) percent security deposit of the net value of the item at the time of delivery of the equipment.
11. In case of refusal/failure to supply, the EMD of the selected tenderer will be forfeited.
12. Packing, forwarding & installation charges should be borne by the supplier. Rate should be inclusive of all.
13. Time of completion of delivery will be directed in order.
14. Service back up of the quoted equipment/instrument should be available in Kolkata.
15. Earnest Money of the unsuccessful bidders will be refunded in due course.
16. The Suppliers should have knowledge to supply the same kind of equipments to other Medical Colleges within the jurisdiction of Kolkata.
17. All suits arising out of the tender/ enquiry and subsequent purchase order/letter of acceptance (if any) are subjected to the jurisdiction in Kolkata only.

Enclosed:

Technical Specifications of "Medical Equipment for the department of Ophthalmology"


Principal
Calcutta National Medical College

Memo No. CNMC/ 286 /P

Date: 02/02/2021

Copy forwarded for information & necessary publication to:

1. The DDHS (E & S), Central Medical Stores,
2. The MSVP, CNMC&H
3. The Accounts Officer, CNMC
4. I.T. Cell, Swasthya Bhavan, for publication in the website
5. Notice Board, CNMC, Kolkata-14
6. CNMC, website


Principal
Calcutta National Medical College

Item No 1

Technical Specification for Applanation Tonometer

Type :
Principle : Applanation Tonometer
Goldmann Tonometry

Specifications

Measuring Range : From 0 to 80mmHg in 2 mmHg increments
Accuracy : +/- 0.5mmHg
Diameter of the pneumatic face : 3.06 mm
Area of Applanation : 7.354 mm²
Probe Carrier Line Separation : 43°
Measurement : 47mm Wide x 30mm Deep x 85mm Height

Standard Accessories

- Calibration Bar
- Prism
- Mount Base (For Head Mount Model)

Compatibility : Both for Zeiss & Haag – Streit Model Slit Lamps.

Am
12/11/2021
HOD of Ophthalmology
Calcutta National Medical College & Hospital
Kolkata-700014

Item No 2

Technical specification for 90 dioptre lens (VOLK/OCULAR)

Make should be VOLK/OCULAR

Item No 3

Technical specification for Two mirror Gonioscope

Make should be VOLK/OCULAR

Item No 4

Technical specification for Zeiss Four mirror gonioscope (VOLK/OCULAR)

Make should be VOLK/OCULAR

G.
14/11/2021
HOD of Ophthalmology
Calcutta National Medical College & Hospital
Kolkata-700014

Item no 5

Technical specification for Reister Sciotz Tonometer

Precision measurement on a scale of 0 to 20 subdivisions and 0 to -1 sub-divisions

Red pointer

Three weights (5.5 g, 7.5 g, 10 g) and a conversion table

Steel body with chrome plating

Proper Casing

Am
12/2014
HOD of Ophthalmology
Calcutta National Medical College & Hospital
Kolkata-700014

Item No 6

Specification for Indirect Ophthalmoscope with 20 Dioptre lens

A) Indirect Ophthalmoscope

Focus distance: At least 300 - 800 mm

Inter Pupillary distance : Minimum 50 to 74 mm

Pupil size optics 1.00 mm

Illumination Halogen illumination

Field : Clear circular

Illumination spot size: $\varnothing 4.0$ mm, $\varnothing 3.0$ mm, $\varnothing 1.2$ mm

Illumination area: $\varnothing 80$ mm, $\varnothing 60$ mm, $\varnothing 25$ mm (at the distance 500mm)

Illumination source: Halogen bulb on the power source

Illumination control: Continuously variable

Filters: Heat absorbing, Green, Cobalt blue

Power supply ~110-230 V, 50-60 Hz

Light source 6 V

6V AC / 2 A power supply unit

Carrying case

Scleral depressor

Fundus chart

Marking pencils

Instruction manual

Double sided Co-observation teaching mirror

B) 20 Diopter Lens (Volk/ Ocular)

Make should be VOLK/OCULAR

12.1.2014
HOD of Ophthalmology
Calcutta National Medical College & Hospital
Kolkata-700014